SIEMENS

MODULARIS Vario

SP

Startup

System

Release for Operation

© Siemens AG 2003
The reproduction, transmission or use of this document or its contents is not permitted without express written authority. Offenders will be liable for damages. All rights, including rights created by patent grant or registration of a utility model or design, are reserved.

Print No.: SPL1-130.815.01.02.02
Replaces: SPL1-130.815.01.01.02

English
Doc. Gen. Date: 08.03

Document revision level

The document corresponds to the version/revision level effective at the time of system delivery. Revisions to hardcopy documentation are not automatically distributed.

Please contact your local Siemens office to order current revision levels.

Disclaimer

The installation and service of equipment described herein is to be performed by qualified personnel who are employed by Siemens or one of its affiliates or who are otherwise authorized by Siemens or one of its affiliates to provide such services.

Assemblers and other persons who are not employed by or otherwise directly affiliated with or authorized by Siemens or one of its affiliates are directed to contact one of the local offices of Siemens or one of its affiliates before attempting installation or service procedures.

Table of Contents 3

1 _____ General Remarks _____ 4

2 _____ Release for Operation _____ 5

3 _____ Changes _____ 6

An individual release is required for each combination of LITHOSTAR MODULARIS Vario with an X-ray C-Arm unit. Prerequisite is a complete startup specific to the Serial Numbers of the units.

If a LITHOSTAR MODULARIS Vario is operated with more than one C-Arm unit, a separate Release Certificate must be generated for each combination.

The following page - Release for Operation - must be sent to:

SP Q Business Unit

FAX: (0049) 9131 84 8750

The response with the signed release must be filed in the System Binder that is with the user. Please file a copy of the Dimensional Protocol along with it in the PG (SPL1-130.891.01).

LITHOSTAR MODULARIS Vario	Siemens 11 58 059, Model J1048
Serial No.:	

in combination with:

X-ray C-Arm Unit	Manufacturer:
Model / Order Code:	
Serial No.:	

Customer:	
Country:	City:

Measure	met
The PG Dimensional Protocol (SPL1-130.891.01..) was available during installation.	
The functionality of the combination was tested prior to startup.	
The mechanical coupling corresponds to the Installation Instructions (SPL1-130.814.01..)	
No additional technical changes have been made to the units.	
The focus adjustment in the crosshairs corresponds to the requirements of the Startup Instructions (SPL1-130.814.01..).	

Name	Date	Signature

Confirmation of Release for Operation

The combination described above is hereby released for shock wave operation. This release applies for the technical version on the day of turnover to the customer. It is no longer valid if changes are made to the technical operating requirements.

SP QP Business Unit

FAX: 049 9131 84 8750

Please use this fax number for all other inquiries regarding release for operation.

Name	Date	Signature

Changes to Previous Version

Layout of complete document.